

# Web Application for the Engineering Data Files Processing

K.Salminsh

Institute of Astronomy  
University of Latvia  
*kalvis@lanet.lv*


# Basic EDF contents

- Epoch (MJD, ISO date)
- Station ID (Name, SOD)
- Link to NP (SCH, SCI, calibration method, time scale)
- Laser (Wavelength, pulse width, rep. rate)
- Timer
- Detector (type, model)
- Calibration data (value, RMS, target distance)
- Meteo data (pressure, humidity, temperature)

# Basic EDF XML

- <EDF Version="1.0" MJD="53744.982639" Epoch="2006-01-09T23:35:00"
- xmlns:Wettzell="http://www.astr.lu.lv/Wettzell">
- <Station SOD="88341001" SCH="0" SCI="2" CalibMethod="1" TimeScale="3" Name="Wettzell"/>
- <Hardware >
- <Laser Wavelength="532.0" Energy="0.075" PulseWidth="60" Divergence="0.000250"
- RepRate="10"/>
- <Receiver>
- <Detector Model="ITT F4129F" DeviceID="1" DetectorType="MCP" TWCompensation="No"/>
- <Filter Model="BARR ASSOCIATES" DeviceID="1002347" BandWidth="0.35"/>
- </Receiver>
- <Timer Model="PET4" DeviceID="Module 1+2+3+4" CorrectionID="0"/>
- </Hardware >
- <Meteo Temperature=" -9.11" Pressure="954.70" Humidity="97.50"/>
- <Calibration TargetDistance="10.345" CalValue="40220" PeakMinusMean="5" RecordedPoints="1200"
- AcceptedPoints="1165" SigmaUsed="2.2" RMS="10" Skew="0.0001" Kurtosis="0.0000"
- Wettzell:ReturnQuote="97.08"/>
- <CustomData> </CustomData >
- </EDF >

# EDF Data Flow



# Application functionality

- Overview of the used equipment
  - Query based interface
  - Map based interface
- Calibration time series
- Calibration charts
- Statistics
- External interfaces (RSS feeds, web services, Excel)

# Query form

The screenshot shows a web browser window with the following content:

- Address bar: <http://127.0.0.1:8084/EXEC/2/0zk05t414y619g1h4b53h17eimd3>
- Form fields:
  - 1-1-2004 (dropdown)
  - 17-10-2006 (dropdown)
  - Total: 39017
  - Timers Used: -- No Selection -- (dropdown), Graz\_ET, PET4
  - Detector models: -- No Selection -- (dropdown), C-SPAD, ITT F4129F
  - Detector types: -- No Selection -- (dropdown), APD, MCP
  - Laser rep. rate: -- No Selection -- (dropdown), 10, 2000
  - Per station: Graz (dropdown)
  - 7534 (text input)
- Buttons: Reload, Back
- Status bar: Done, Trusted sites, 100%

# Calibration data

Start: 1-1-2006 Stop: 19-10-2006  
Wetzell

< 61 62 63 64 65 66 67 68 69 70 > 61 of 2020

EPOCH	DETECTOR_TYPE	DETECTOR_MODEL	CAL_VALUE	RMS	TEMPERATURE	HUMIDITY	PRESSURE
2006-04-23 17:20:15	APD	Silicon Sensor	42225	71	14.69	64.8	942.53
2006-04-23 17:23:47	APD	Silicon Sensor	42211	74	14.69	64.8	942.53
2006-04-23 17:27:06	APD	Silicon Sensor	42220	65	14.69	64.8	942.53
2006-04-23 17:30:31	APD	Silicon Sensor	42220	68	14.69	64.8	942.53
2006-04-23 17:41:16	MCP	ITT F4129F	40195	11	13.99	61	942.92
2006-04-23 17:43:18	MCP	ITT F4129F	40190	11	13.99	61	942.92
2006-04-23 17:45:18	MCP	ITT F4129F	40190	11	13.99	61	942.92
2006-04-23 17:47:18	MCP	ITT F4129F	40188	7	13.99	61	942.92
2006-04-23 18:24:40	MCP	ITT F4129F	40196	12	13.19	63.3	943.12
2006-04-23 18:26:40	MCP	ITT F4129F	40192	11	13.19	63.3	943.12

[Basic calibr. data](#)

# Next steps

- Integrate application in the EDF website  
<http://www.astr.lu.lv/EDF>
- User interface improvements
- Establish links to ILRS website
- Move to native XML database
- Links with the NP, site logs